

X10C SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: C	Effective: 5/24/2010	Page 1 of 1	Number: LS-SCL-0023
Prepared by: Philip Marino		Reviewed by: P. Marino	Approved by: B. Chmiel

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

- ☐ 1. Valid Padlock Index and log, no padlocks open

Checkpoints on the X9 side of beamline:

- ☐ 2. Bremsstrahlung shield (BS # 1) in place and banded as per photo
- ☐ 3. Bremsstrahlung shield (BS # 2) in place and banded as per photo
- ☐ 4. Exclusion Zone (EZ #1) wire mesh in place and secure
- ☐ 5. Bremsstrahlung shield (BS # 3) in place and banded as per photo
- ☐ 6. Slit window covered with lead
- ☐ 7. Bremsstrahlung shield (BS # 4) in place and banded as per photo
- ☐ 8. Monochromator window shielded with leaded glass

Checkpoints on the X11 side of beamline:

- ☐ 9. Monochromator window shielded with leaded glass
- ☐ 10. Bremsstrahlung shield (BS # 5) in place and banded as per photo
- ☐ 11. Mirror windows (3) covered in lead as per photo

OPCO / Beamline Staff: _____ Date: _____

USER SYSTEM CHECKS

- ☐ A. Beamline Isolation Valve A is open
- ☐ B. Beamline Isolation Valve B is open
- ☐ C. Beamline Isolation Valve C is open
- ☐ D. Beamline Isolation Valve D is open

Beamline Rep.: _____ Date: _____